

## Data sheet Monkey GM-CSF ELISPOT antibody pair; 10-plate format

U-CyTech BV Yalelaan 48 3584 CM Utrecht The Netherlands P +31.30.253 5960 F +31.30.253 9344 INFO@ucytech.com

Cat. No.:

Coating antibodies (2 vials)	
Product:	Monoclonal antibody to monkey granulocyte-macrophage colony stimulating factor (GM-CSF)
lsotype:	Mouse IgG <sub>1</sub>
Production:	In vitro using serum free medium
Purification:	DEAE ion exchange chromatography
Contents:	Each vial contains sufficient material for coating of five 96-well ELISPOT plates
Buffer:	Prior to lyophilization: 0.25 ml PBS + 125 mM trehalose
Application:	Coating antibody in an ELISPOT system
Reconstitution:	Dissolve the contents of one vial by injection of 0.25 ml distilled water into the vial and dilute 100 times in PBS. The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; $50 \mu$ l/well).

## Detection antibodies (2 vials)

Product: Biotinylated monoclonal antibody to monkey granulocyte-macrophage colony stimulating factor (GM-CSF)

Isotype: Rat IgG<sub>2a</sub>

Production: In vitro using serum free medium

CT608-10

Purification: Protein G-affinity chromatography

Labeling: With Biotin-7-NHS (N-hydroxysuccinimide)

Contents: Each vial contains sufficient material for five 96-well ELISPOT plates

Buffer: Prior to lyophilization: 0.5 ml PBS + 1% BSA + 125 mM trehalose

Application: Detection antibody in an ELISPOT system

Reconstitution: Dissolve the contents of one vial by injection of 0.5 ml distilled water into the vial and dilute 100 times in Dilution buffer (see Technical Data Sheet). The total amount of one vial is sufficient for five 96-well ELISPOT plates (480 determinations; 100  $\mu$ I/well).

## General

Specificity: Validated for detecting rhesus macaque, cynomolgus monkey and mandrill GM-CSF

Sterility: Membrane filtered (0.2 µm)

Stability: The lyophilized products are stable for at least one year at 4°C (expiry date is indicated on the vials).
After reconstitution, the antibodies are stable for several months at 4°C (if kept sterile) or for minimal one year at -20°C.

